

AMS S.p.A.  
Via E.Barsanti 17/A  
Guidonia, ROME  
I-00012

Phone : +39 0774 354441 Fax : +39 0774 578035

Email : info@ams-analyzers.com Web : www.ams-analyzers.com

Method : CR6+ -Unit [ mg/L ] - CHROMIUM VI (0-0.25 PPM)

Smp#[/Dil Fact]	Sample ID	Conc.	OD	%Recovery/RPD	Flag	Analysis Time
DIL-1	RBL	0.0000	0.0023	0.00		11:29:54 AM
DIL-1	RBL	0.0000	0.0019	0.00		11:30:12 AM
DIL-1	Std-1	0.0000	0.0000	0.00		11:30:30 AM
SR5-1	Std-2	0.0250	0.0064	0.00		11:30:48 AM
SR5-2	Std-3	0.0500	0.0172	0.00		11:31:06 AM
SR5-3	Std-4	0.1000	0.0352	0.00		11:31:24 AM
SR5-4	Std-5	0.1563	0.0556	0.00		11:31:42 AM
SR5-5	Std-6	0.2500	0.0874	0.00		11:32:00 AM
1	Blk DI -1	0.0028	0.0002	0.00		11:34:42 AM
2	2	0.0014	-0.0003	0.00	INV,><	11:35:00 AM
3	3	0.0022	0.0000	0.00		11:35:18 AM
4	4	-0.0017	-0.0014	0.00	INV,><,LL	11:35:36 AM
5	5	0.0031	0.0003	0.00		11:35:54 AM
6	6	-0.0006	-0.0010	0.00	INV,><,LL	11:36:12 AM
7	7	0.0014	-0.0003	0.00	INV,><	11:36:31 AM
8	8	0.0022	0.0000	0.00		11:36:48 AM
9	9	0.0020	-0.0001	0.00	INV,><	11:39:30 AM
10	River Water Blk 1	0.0118	0.0034	0.00		11:39:48 AM
11	-2	0.0118	0.0034	0.00		11:40:06 AM
12	-3	0.0124	0.0036	0.00		11:40:24 AM
13	-4	0.0124	0.0036	0.00		11:40:43 AM
14	-5	0.0127	0.0037	0.00		11:41:01 AM
15	-6	0.0129	0.0038	0.00		11:41:19 AM
16	-7	0.0124	0.0036	0.00		11:41:37 AM
17	-8	0.0127	0.0037	0.00		11:44:19 AM
18	-9	0.0124	0.0036	0.00		11:44:37 AM
19	QC1	0.0278	0.0091	0.00		11:44:55 AM
20	Blk 1	0.0025	0.0001	0.00		11:45:13 AM
21	RW 0.005 ppm -1	0.0174	0.0054	0.00		11:45:31 AM
22	2-	0.0183	0.0057	0.00		11:45:49 AM
23	3-	0.0174	0.0054	0.00		11:46:07 AM
24	4-	0.0172	0.0053	0.00		11:46:25 AM
25	5-	0.0180	0.0056	0.00		11:49:07 AM
26	6-	0.0177	0.0055	0.00		11:49:25 AM
27	7-	0.0183	0.0057	0.00		11:49:43 AM
28	8-	0.0180	0.0056	0.00		11:50:01 AM

Report Date :01/18/2013

Operator :WESTCO

Plan Description : Cr 6 + in River Water

Spiked 0.005 mg/L, 0.05 mg/L and 0.25 mg/L in River Water