

AMS S.p.A.

Via E.Barsanti 17/A

Guidonia, ROME

I-00012

Phone : +39 0774 354441 Fax : +39 0774 578035

Email : info@ams-analyzers.com Web : www.ams-analyzers.com

Method : CR6+ -Unit [mg/L] - CHROMIUM VI (0-0.25 PPM)

Smp#[/Dil Fact]	Sample ID	Conc.	OD	%Recovery/RPD Flag	Analysis Time
29	9-	0.0177	0.0055	0.00	11:50:19 AM
30	QC 2	0.0284	0.0093	0.00	11:50:37 AM
31	Blk 2	0.0022	0.0000	0.00	11:50:55 AM
32	RW 0.05 ppm -1	0.0605	0.0207	0.00	11:51:13 AM
33	--2	0.0611	0.0209	0.00	11:53:55 AM
34	--3	0.0619	0.0212	0.00	11:54:13 AM
35	--4	0.0616	0.0211	0.00	11:54:31 AM
36	--5	0.0611	0.0209	0.00	11:54:49 AM
37	--6	0.0613	0.0210	0.00	11:55:07 AM
38	--7	0.0613	0.0210	0.00	11:55:25 AM
39	--8	0.0616	0.0211	0.00	11:55:43 AM
40	--9	0.0616	0.0211	0.00	11:56:01 AM
41	QC 3	0.0284	0.0093	0.00	11:58:43 AM
42	Blk 3	0.0022	0.0000	0.00	11:59:01 AM
43	RW 0.25 ppm -1	0.2471	0.0870	0.00	11:59:19 AM
44	2---	0.2456	0.0865	0.00	11:59:37 AM
45	3---	0.2462	0.0867	0.00	11:59:55 AM
46	4---	0.2440	0.0859	0.00	12:00:13 PM
47	5---	0.2448	0.0862	0.00	12:00:31 PM
48	6---	0.2442	0.0860	0.00	12:00:49 PM
49	7---	0.2487	0.0876	0.00 ><	12:03:31 PM
50	8---	0.2479	0.0873	0.00	12:03:49 PM
51	9---	0.2473	0.0871	0.00	12:04:07 PM
52	QC 4	0.0281	0.0092	0.00	12:04:25 PM
53	Blk 4	0.0017	-0.0002	0.00 INV,><	12:04:43 PM

Report Date :01/18/2013

Operator : WESTCO

Plan Description : Cr 6 + in River Water

Spiked 0.005 mg/L, 0.05 mg/L and 0.25 mg/L in River Water