

AMS S.p.A.
Via E.Barsanti 17/A
Guidonia, ROME
I-00012

Phone : +39 0774 354441 Fax : +39 0774 578035

Email : info@ams-analyzers.com Web : www.ams-analyzers.com

Method: CR6+ -Unit [mg/L] - CHROMIUM VI (0-0.25 PPM)

Smp#[/Dil Fact]	Sample ID	Conc.	OD	%Recovery/RPD	Flag	Analysis Time
DIL-1	RBL	0.0000	0.0015	0.00		11:39:28 AM
DIL-1	RBL	0.0000	0.0013	0.00		11:39:46 AM
DIL-1	Std-1	0.0000	-0.0002	0.00	INV	11:40:04 AM
SR5-1	Std-2	0.0250	0.0124	0.00		11:40:22 AM
SR5-2	Std-3	0.0500	0.0244	0.00		11:40:40 AM
SR5-3	Std-4	0.1000	0.0483	0.00		11:40:58 AM
SR5-4	Std-5	0.1563	0.0780	0.00		11:41:16 AM
SR5-5	Std-6	0.2500	0.1224	0.00		11:41:34 AM
1	Blk 1	0.0002	0.0000	0.00		11:44:16 AM
2	2	0.0002	0.0000	0.00		11:44:34 AM
3	3	0.0004	0.0001	0.00		11:44:52 AM
4	4	0.0006	0.0002	0.00		11:45:10 AM
5	5	0.0004	0.0001	0.00		11:45:28 AM
6	6	0.0002	0.0000	0.00		11:45:46 AM
7	7	0.0000	-0.0001	0.00	INV	11:46:05 AM
8	8	0.0002	0.0000	0.00		11:46:23 AM
9	9	0.0002	0.0000	0.00		11:49:05 AM
10	10	0.0000	-0.0001	0.00	INV	11:49:22 AM
11	SW blank spl -1	0.0094	0.0045	0.00		11:49:40 AM
12	2-	0.0023	0.0010	0.00		11:49:59 AM
13	3-	0.0023	0.0010	0.00		11:50:17 AM
14	4-	0.0021	0.0009	0.00		11:50:35 AM
15	5-	0.0019	0.0008	0.00		11:50:53 AM
16	6-	0.0019	0.0008	0.00		11:51:11 AM
17	7-	0.0025	0.0011	0.00		11:53:53 AM
18	8-	0.0025	0.0011	0.00		11:54:11 AM
19	9-	0.0021	0.0009	0.00		11:54:29 AM
20	10-	0.0025	0.0011	0.00		11:54:47 AM
21	QC 0.05 ppm 1	0.0515	0.0252	0.00		11:55:05 AM
22	SW	0.0031	0.0014	0.00		11:55:23 AM
23	SW spike 0.005ppm-1	0.0073	0.0035	0.00		11:55:41 AM
24	2--	0.0071	0.0034	0.00		11:55:59 AM
25	3--	0.0100	0.0048	0.00		11:58:41 AM
26	4--	0.0075	0.0036	0.00		11:58:59 AM
27	5--	0.0067	0.0032	0.00		11:59:17 AM
28	6--	0.0077	0.0037	0.00		11:59:35 AM

Report Date :01/18/2013

Operator : WESTCO

Plan Description : Cr 6+ in SW

Spiked 0.005 mg/L, 0.05 mg/L and 0.25 mg/L in Seawater