


Calibrant Report - CR6+ -

Calib Lot #:1
Plan #: 20110811005

Spiked 0.05 mg/L and 0.25 mg/L in Deionized Water



Point	OD	Conc.	Recalc Conc	% Error
1	0.0012	0	0.0010	0.10
2	0.0119	0.025	0.0222	-11.20
3	0.0259	0.05	0.0500	0.00
4	0.0514	0.1	0.1007	0.70
5	0.0806	0.1563	0.1587	1.54
6	0.1258	0.25	0.2485	-0.60

Conc. = +1.9862*Abso -0.0014 R²=0.9996

RBL
-0.0012
0