

AMS S.p.A.
Via E.Barsanti 17/A
Guidonia, ROME
I-00012

Phone : +39 0774 354441 Fax : +39 0774 578035

Email : info@ams-analyzers.com Web : www.ams-analyzers.com

Method : CR6+ -Unit [mg/L] - CHROMIUM VI (0-0.25 PPM)

Smp#[/Dil Fact]	Sample ID	Conc.	OD	%Recovery/RPD	Flag	Analysis Time
DIL-1	RBL	0.0000	-0.0016	0.00	R	3:59:13 PM
DIL-1	RBL	0.0000	-0.0007	0.00	R	3:59:31 PM
DIL-1	Std-1	0.0000	0.0013	0.00		3:59:49 PM
SR5-1	Std-2	0.0250	0.0119	0.00		4:00:07 PM
SR5-2	Std-3	0.0500	0.0260	0.00		4:00:25 PM
SR5-3	Std-4	0.1000	0.0515	0.00		4:00:43 PM
SR5-4	Std-5	0.1563	0.0806	0.00		4:01:01 PM
SR5-5	Std-6	0.2500	0.1258	0.00		4:01:19 PM
1	Blk 1	0.0015	0.0015	0.00		4:04:01 PM
2	2	0.0017	0.0016	0.00		4:04:19 PM
3	3	0.0023	0.0019	0.00		4:04:37 PM
4	4	0.0019	0.0017	0.00		4:04:55 PM
5	5	0.0019	0.0017	0.00		4:05:13 PM
6	6	0.0023	0.0019	0.00		4:05:31 PM
7	7	0.0021	0.0018	0.00		4:05:49 PM
8	8	0.0021	0.0018	0.00		4:06:08 PM
9	9	0.0021	0.0018	0.00		4:08:49 PM
10	10	0.0023	0.0019	0.00		4:09:08 PM
11	0.05-1	0.0522	0.0270	0.00		4:09:25 PM
12	0.05-2	0.0530	0.0274	0.00		4:09:43 PM
13	0.05-3	0.0522	0.0270	0.00		4:10:02 PM
14	0.05-4	0.0524	0.0271	0.00		4:10:20 PM
15	0.05-5	0.0514	0.0266	0.00		4:10:38 PM
16	0.05-6	0.0524	0.0271	0.00		4:10:56 PM
17	0.05-7	0.0504	0.0261	0.00		4:13:38 PM
18	0.05-8	0.0520	0.0269	0.00		4:13:56 PM
19	0.05-9	0.0510	0.0264	0.00		4:14:14 PM
20	0.05-10	0.0522	0.0270	0.00		4:14:32 PM
21	QC 5ppm	0.0530	0.0274	0.00		4:14:50 PM
22	Blk	0.0011	0.0013	0.00		4:15:08 PM
23	0.10-1	0.1012	0.0517	0.00		4:15:26 PM
24	0.10-2	0.1020	0.0521	0.00		4:15:44 PM
25	0.10-3	0.1028	0.0525	0.00		4:18:26 PM
26	0.10-4	0.1020	0.0521	0.00		4:18:44 PM
27	0.10-5	0.1024	0.0523	0.00		4:19:02 PM
28	0.10-6	0.1032	0.0527	0.00		4:19:20 PM

Report Date :01/18/2013

Operator : WESTCO

Plan Description : Cr 6+

Spiked 0.05 mg/L and 0.25 mg/L in Deionized Water