

AMS S.p.A.

Via E.Barsanti 17/A

Guidonia, ROME

I-00012

Phone : +39 0774 354441 Fax : +39 0774 578035

Email : info@ams-analyzers.com Web : www.ams-analyzers.com

Method : CR6+ -Unit [mg/L] - CHROMIUM VI (0-0.25 PPM)

Smp#[[Dil Fact]	Sample ID	Conc.	OD	%Recovery/RPD	Flag	Analysis Time
29	0.10-7	0.1024	0.0523	0.00		4:19:38 PM
30	0.10-8	0.1016	0.0519	0.00		4:19:56 PM
31	0.10-9	0.1004	0.0513	0.00		4:20:14 PM
32	0.10-10	0.1018	0.0520	0.00		4:20:32 PM
33	QC	0.0536	0.0277	0.00		4:23:14 PM
34	Blk -	0.0019	0.0017	0.00		4:23:32 PM
35	0.25-1	0.2518	0.1275	0.00	><,LH	4:23:50 PM
36	0.25-2	0.2512	0.1272	0.00	><,LH	4:24:08 PM
37	0.25-3	0.2516	0.1274	0.00	><,LH	4:24:26 PM
38	0.25-4	0.2514	0.1273	0.00	><,LH	4:24:44 PM
39	0.25-5	0.2502	0.1267	0.00	><,LH	4:25:02 PM
40	0.25-6	0.2508	0.1270	0.00	><,LH	4:25:20 PM
41	0.25-7	0.2512	0.1272	0.00	><,LH	4:28:02 PM
42	0.25-8	0.2514	0.1273	0.00	><,LH	4:28:20 PM
43	0.25-9	0.2536	0.1284	0.00	><,LH	4:28:38 PM
44	0.25-10	0.2522	0.1277	0.00	><,LH	4:28:56 PM
45	QC 1	0.9499	0.4790	0.00	><,LH	4:29:14 PM
46	1 ppm	0.0534	0.0276	0.00		4:29:32 PM

Report Date :01/18/2013

Operator : WESTCO

Plan Description : Cr 6+

Spiked 0.05 mg/L and 0.25 mg/L in Deionized Water