

AMS S.p.A.
Via E.Barsanti 17/A
Guidonia, ROME
I-00012

Phone : +39 0774 354441 Fax : +39 0774 578035

Email : info@ams-analyzers.com Web : www.ams-analyzers.com

Method : CR6+ -Unit [mg/L] - CHROMIUM VI (0-0.25 PPM)

| Smp#[/Dil Fact] | Sample ID | Conc. | OD | %Recovery/RPD | Flag | Analysis Time |
|-----------------|-------------|---------|---------|---------------|-------|---------------|
| DIL-1 | RBL | 0.0000 | -0.0014 | 0.00 | R | 11:49:47 AM |
| DIL-1 | RBL | 0.0000 | -0.0008 | 0.00 | R | 11:50:05 AM |
| DIL-1 | Std-1 | 0.0000 | 0.0010 | 0.00 | | 11:50:23 AM |
| SR5-1 | Std-2 | 0.0250 | 0.0131 | 0.00 | | 11:50:41 AM |
| SR5-2 | Std-3 | 0.0500 | 0.0258 | 0.00 | | 11:50:59 AM |
| SR5-3 | Std-4 | 0.1000 | 0.0509 | 0.00 | | 11:51:17 AM |
| SR5-4 | Std-5 | 0.1563 | 0.0783 | 0.00 | | 11:51:35 AM |
| SR5-5 | Std-6 | 0.2500 | 0.1255 | 0.00 | | 11:51:53 AM |
| 1 | Blk 1 | 0.0001 | 0.0009 | 0.00 | >< | 11:54:35 AM |
| 2 | 2 | 0.0005 | 0.0011 | 0.00 | | 11:54:53 AM |
| 3 | 3 | 0.0009 | 0.0013 | 0.00 | | 11:55:11 AM |
| 4 | 4 | 0.0005 | 0.0011 | 0.00 | | 11:55:29 AM |
| 5 | 5 | 0.0001 | 0.0009 | 0.00 | >< | 11:55:47 AM |
| 6 | 6 | 0.0009 | 0.0013 | 0.00 | | 11:56:05 AM |
| 7 | 7 | 0.0011 | 0.0014 | 0.00 | | 11:56:23 AM |
| 8 | 8 | 0.0007 | 0.0012 | 0.00 | | 11:56:41 AM |
| 9 | 9 | 0.0009 | 0.0013 | 0.00 | | 11:59:23 AM |
| 10 | 10 | 0.0009 | 0.0013 | 0.00 | | 11:59:41 AM |
| 11 | 0.005 ppm-1 | 0.0065 | 0.0041 | 0.00 | | 11:59:59 AM |
| 12 | a-2 | 0.0057 | 0.0037 | 0.00 | | 12:00:17 PM |
| 13 | a-3 | 0.0061 | 0.0039 | 0.00 | | 12:00:35 PM |
| 14 | a-4 | 0.0057 | 0.0037 | 0.00 | | 12:00:53 PM |
| 15 | a-5 | 0.0047 | 0.0032 | 0.00 | | 12:01:12 PM |
| 16 | a-6 | 0.0063 | 0.0040 | 0.00 | | 12:01:30 PM |
| 17 | a-7 | 0.0049 | 0.0033 | 0.00 | | 12:04:12 PM |
| 18 | a-8 | 0.0061 | 0.0039 | 0.00 | | 12:04:30 PM |
| 19 | a-9 | 0.0055 | 0.0036 | 0.00 | | 12:04:48 PM |
| 20 | a-10 | 0.0061 | 0.0039 | 0.00 | | 12:05:06 PM |
| 21 | QC 5ppm | 0.0531 | 0.0273 | 0.00 | | 12:05:24 PM |
| 22 | Blk-1 | -0.0009 | 0.0004 | 0.00 | ><,LL | 12:05:42 PM |
| 23 | 0.01 ppm-1 | 0.0107 | 0.0062 | 0.00 | | 12:06:00 PM |
| 24 | b-2 | 0.0117 | 0.0067 | 0.00 | | 12:06:18 PM |
| 25 | b-3 | 0.0113 | 0.0065 | 0.00 | | 12:09:00 PM |
| 26 | b-4 | 0.0099 | 0.0058 | 0.00 | | 12:09:18 PM |
| 27 | b-5 | 0.0121 | 0.0069 | 0.00 | | 12:09:36 PM |
| 28 | b-6 | 0.0115 | 0.0066 | 0.00 | | 12:09:54 PM |

Report Date :01/18/2013

Operator : WESTCO

Plan Description : Cr 6 +

Spiked 0.005 mg/L in Deionized Water