


Calibrant Report - CR6+ -

Calib Lot #:1

Spiked 0.005 mg/L, 0.05 mg/L and 0.25 mg/L in Seawater

Plan # : 20110812001



Point	OD	Conc.	Recalc Conc	% Error
1	-0.0002	0	-0.0002	-0.02
2	0.0124	0.025	0.0254	1.60
3	0.0244	0.05	0.0498	-0.40
4	0.0483	0.1	0.0984	-1.60
5	0.0780	0.1563	0.1588	1.60
6	0.1224	0.25	0.2490	-0.40

Conc. = +2.0328*Abso +0.0002 R²=0.9998

RBL
0.0014
0