


# Calibrant Report - CR6+ -

Calib Lot #:1

Spiked 0.005 mg/L, 0.05 mg/L and 0.25 mg/L in Industrial Effluent

Plan # : 20110816002



Point	OD	Conc.	Recalc Conc	% Error
1	0.0000	0	-0.0002	-0.02
2	0.0091	0.025	0.0237	-5.20
3	0.0190	0.05	0.0497	-0.60
4	0.0388	0.1	0.1017	1.70
5	0.0604	0.1563	0.1584	1.34
6	0.0947	0.25	0.2484	-0.64

Conc. = +2.6253\*Abso -0.0002    R<sup>2</sup>=0.9997

RBL  
0.0021  
0